

# Manual Supplement

Manual Title: 725Ex Users  
Print Date: January 2005  
Revision/Date: Rev.2, 5/09

Supplement Issue: **4**  
Issue Date: 11/11  
Page Count: 2

---

---

This supplement contains information necessary to ensure the accuracy of the above manual. This manual is distributed as an electronic manual on the following CD-ROM:

CD Title:	725Ex Users
CD Rev. & Date:	2, 5/2009
CD PN:	2406548

## Change #1, 52976

On pages 10 and 59, under ***Certification Information & General Specification, Product Compliance Markings***, change the following:

From: CE                       II 1 G EEx ia IIB 171 °C  
0344                              KEMA 04ATEX1303

To:                              CE                               II 1 G Ex ia IIB 171 °C  
0344                              KEMA 04ATEX1303

## Change #2, 54531

On page 10, under ***Certification Information***, replace the fourth bullet with the following:

- Manufactured by Martel Electronics, Inc.  
                            3 Corporate Park Drive  
                            Derry, NH USA

## Change #3

On page 52, Table 9:

Change:

Test lead, red	688051	1
Test lead, black	688066	1

To:

Fluke-7XX Test Lead Set	3397308	1
-------------------------	---------	---

## Change #4

On page 6, add the following Caution:

### Caution

**The 725Ex MEASURE/SOURCE terminals are ESD (electro-static discharge) sensitive to levels above  $\pm 2$  kV. The Calibrator can experience temporary loss of measurement or source functionality, which may require operator intervention to restore product function, or even cause permanent damage to the product. In general, a disruptive ESD event will only occur during connection of the test leads to the circuits being measured or if the operator is carrying a large static charge and touches the Calibrator terminals. The most common cause of ESD is the user carrying the Calibrator across a carpet, or other similar triboelectric activity, before they connect to the circuit being measured.**