

Manual Supplement

Manual Title: 1623-2 Users
Print Date: January 2014
Revision/Date:

Supplement Issue: 1
Issue Date: 8/14
Page Count: 1

This supplement contains information necessary to ensure the accuracy of the above manual. This manual is distributed as an electronic manual on the following CD-ROM:

CD Title: 1623-2/1625-2
CD Rev. & Date: 1/2014
CD PN: 4325414

Change #1

On page 30, replace the table with:

Resolution	Range	Accuracy	Operating Error
0.001	0.020 to 1.000 Ω	\pm (2 % rdg + 10 digits)	\pm (5 % rdg + 3 digits)
0.001 to 10 Ω	1.001 to 19.99 k Ω	\pm (2 % rdg +3 digits)	\pm (5 % rdg + 3 digits)

On page 31, replace both tables with:

Resolution	Range	Accuracy	Operating Error
0.001	0.020 to 1.000 Ω	\pm (7 % rdg + 10 digits)	\pm (10 % rdg + 5 digits)
0.001 to 10 Ω	1.001 to 19.99 k Ω	\pm (7 % rdg + 3 digits)	\pm (10 % rdg + 5 digits)

Resolution	Range	Accuracy	Operating Error
0.001	0.020 to 1.000 Ω	\pm (7 % rdg + 10 digits)	\pm (10 % rdg + 5 digits)
0.001 to 10 Ω	1.001 to 19.99 k Ω	\pm (7 % rdg + 3 digits)	\pm (10 % rdg + 5 digits)